

ET Enterprises

Ref No. 13B
Order No. 45285



IMPORTANT - This test ticket must be returned with the tube if making a guarantee claim

Type: 9823QKB04
Serial: 16747

TBST RESULT DATA

CB 11.9 CR 2.6 D1 0.
VOLTS @ 5000A/1m 2416.000 V
D/C @ 5000A/1m 72.500 nA
VOLTS @ 10000A/1m 2526.000 V

**Installed,
Top Right
JLAB \$**

form F/DP006 issue 2
2.10.08

ET Enterprises

Ref No. 13B
Order No. 44733



IMPORTANT - This test ticket must be returned with the tube if making a guarantee claim

Type: 9823QKB04
Serial: 16777

TBST RESULT DATA

CB 12.1 CR 2.9 D1 0.8
VOLTS @ 5000A/1m 2493.000 V
D/C @ 5000A/1m 48.000 nA
VOLTS @ 10000A/1m 2637.000 V

**Installed,
Top Left,
UVA \$**

form F/DP006 issue 2
2.10.08

ET Enterprises

Ref No. 13B
Order No. 45285



IMPORTANT - This test ticket must be returned with the tube if making a guarantee claim

Type: 9823QKB04
Serial: 16787

TBST RESULT DATA

CB 8.8 CR 0.2 D1 0.7
VOLTS @ 5000A/1m 2442.000 V
D/C @ 5000A/1m 66.000 nA
VOLTS @ 10000A/1m 2570.000 V

**Installed,
Bottom Right
JLAB \$**

form F/DP006 issue 2
2.10.08

ET Enterprises

Ref No. 13B
Order No. 45285



IMPORTANT - This test ticket must be returned with the tube if making a guarantee claim

Type: 9823QKB04
Serial: 16782

TBST RESULT DATA

CB 9.7 CR 0.6 D1 0.8
VOLTS @ 5000A/1m 2430.000 V
D/C @ 5000A/1m 120.000 nA
VOLTS @ 10000A/1m 2546.000 V

**Died, at UVA
JLAB \$**

form F/DP006 issue 2
2.10.08

ET Enterprises

Ref No. 13B
Order No. 45285



IMPORTANT - This test ticket must be returned with the tube if making a guarantee claim

Type: 9823QKB04
Serial: 16785

TBST RESULT DATA

CB 10.7 CR 1.1 D1 0.7
VOLTS @ 5000A/1m 2260.000 V
D/C @ 5000A/1m 64.900 nA
VOLTS @ 10000A/1m 2366.000 V

**Installed,
Bottom Left,
JLAB \$**

form F/DP006 issue 2
2.10.08

ET Enterprises

Ref No. 13B
Order No. 45285



IMPORTANT - This test ticket must be returned with the tube if making a guarantee claim

Type: 9823QKB04
Serial: 16786

TBST RESULT DATA

CB 7.5 CR 0.8 D1 0.7
VOLTS @ 5000A/1m 2330.000 V
D/C @ 5000A/1m 43.000 nA
VOLTS @ 10000A/1m 2408.000 V

**At UVA 4/6/13,
JLAB \$**

form F/DP006 issue 2
2.10.08