



2016-07-21 MAGNET CONTROL REPORT

Customer : JLAB
Date: 21/07/2016
 $\Sigma\Phi$ reference : 317111
Revision : A
Author: DRA

Evolutions	Date
Création	21/07/2016

Designation :
DIPOLE Magnet

	UNIT	REQUIRED	RESULT	Date Initial
DC resistance value at 20° C	m Ω		2790	15/09/2016 AJ
Dielectric test during 60s at 1500V between coil and ground		OK	OK i=33.9nA	15/09/2016 AJ
Insulation Resistance during 60s at 1500V between coil and ground	G Ω		45	15/09/2016 AJ