


SIGMAPHI REFERENCE 317111
 Serial N° : 1
 Date of control 29/06/2015 by ST & F.L
 & operators:

Polarity control

View from electric side:



Magnetic control @50A around the centre of the magnet

Device used: METROLAB THM1176 Hall probe 3
 axis
 X=Y=Z=0mm

Current approx value (A)	Field modul max value (G)
35	349.93
40	399.34
45	440.5
50	514.64

Expected value for magnetic field @50A = 503 Gauss (without warm iron)